MAC* 2000

Product datasheet



Instrument Type

Microprocessor augmented automatic electrocardiograph; 10-leadwire, 12-lead simultaneous acquisition with programmable lead configuration.

| Processing | |
|-------------------------------|---|
| ECG Interpretation | Marquette* 12SL ECG Analysis Program for Adults and Pediatrics |
| Computerized measurements | 12-lead analysis |
| ECG analysis frequency | 500 or 1000 samples/second/channel |
| Digital sampling rate | 16000 samples/second/channel for normal data acquisition |
| Pace sampling rate | 75K samples/second/channel |
| ECG on-screen preview | On-screen preview of acquired 10 second ECG waveform |
| Acquisition mode | Provides 10 seconds of instantaneous ECG acquisition |
| Dynamic range | AC Differential ±5 mV, DC offset ±300 mV |
| Resolution | 4.88 μV +/-1% per LSB @ 500 SPS |
| Frequency range | 0.04 to 150 Hz |
| Low cut off frequency | 0.04 Hz (ADS off), 0.56 Hz (ADS On) |
| High cut off frequency | Configurable at 20 Hz, 40 Hz, 100 Hz or 150 Hz |
| Common mode rejection | >135 dB (with 50/60 Hz filter ON) |
| Input impedance | >10MΩ @ 10 Hz |
| Patient leakage | <10μA (Normal Condition), <50μA (Single Fault Condition) |
| Lead detection | All disconnected lead detection except RL & RA |
| Heart rate meter | 30 to 300 BPM |
| Operating system | Microsoft® Windows® CE 6.0 |
| Start-up time | Less than 30 seconds |
| Patient Information | |
| Supported patient information | Patient ID, secondary patient ID, last name, first name, height, weight, gender, race, pacemaker patient, systolic BP, diastolic BP, location number, room, order number, phone number, medication, ordering physician, referring physician, attending physician, technician, test indication |



| Display | |
|------------------------------|--|
| Display type | 7 in. color TFT display with support of minimum 32K colors |
| Display resolution | WVGA resolution – 800 × 480 |
| Display data | Heart rate, patient ID, clock, battery power indicator, waveforms, lead lables, speed, gain and filter settings, warning messages, prompts, help messages, and 12-lead display |
| Writer | |
| Writer technology | Thermal dot array |
| Writer speed | 5, 12.5, 25, and 50 mm/s |
| Number of traces | Up to 12 ECG traces |
| Writer sensitivity/gain | 2.5, 5, 10, 20, 40 mm/mV |
| Writer speed accuracy | 5, 12.5 mm/s @ ±5% and 25, 50 mm/s @ ±2% |
| Writer amplitude accuracy | ±5% |
| Writer resolution | Horizontal 40 dots/mm @ 25 mm/s, 8 dots/mm vertical |
| Paper type | Z-fold Thermal Paper with pre-printed grid and perforation with Queue mark or Queue hole |
| Paper size | 215 mm \times 280 mm (letter) and 210 mm \times 295 mm (A4), and 214.2 mm \times 279.4 mm (Modified Letter) |
| Keyboard | |
| Туре | Membrane keyboard with tactile feedback – Soft function keys, alphanumeric keys (Qwerty key set), writer controls and Trim Pad cursor controls |
| Operating modes and additio | nal features |
| Resting ECG mode | Records and prints 12-lead resting ECGs with 10 seconds duration as a standard feature |
| Arrhythmia mode | Continuously monitors ECG and prints report when arrhythmia events of the user-selected class occur |
| Exercise mode | Exercise mode for exercise stress testing |
| RR Analysis mode | RR analysis for RR intervals analysis |
| Hookup advisor | Provides visual indication of signal quality |
| Multi-language support | Supports 19 languages in User Interface and 31 languages in User Manual |
| Order manager | Provides an interface for managing orders |
| ADT Query | Provides an interface for querying patient demographics |
| File manager | Provides an interface for managing ECG Records |
| System setup | Provides an interface for managing Device Configuration |
| Stress/Pharma Application Op | ptions |
| Stress testing application | Ergometers supported include: eBike Treadmills supported include: T2100, T2000 Master's Step device without interface (acoustic signal only) |
| | Note: Ergometer, Master Step, and Treadmill sold separately |
| Pharma application | Pharma application options include: |
| | Audit trail export |
| | • CT Data Guard® |
| | High security login protection |
| External Peripherals | |
| Keyboard | Standard USB English Keyboard |
| Barcode reader | Jadak-1799 |
| | |

| Communication | | | |
|--|---|--|--|
| • Internal modem, Secured Digital card, Serial, LAN, and WIFI communicates outbound to MUSE* and CardioSoft* | | | |
| Internal modem, LAN, and WIFI communicates inbound to MUSE | | | |
| RS232 serial cable | ECG Transmission with A5 & CSI Protocol | | |
| Internal Modem | ECG Transmission with CSI Protocol | | |
| Supported MUSE/CardioSoft | Compatible with MUSE V7.1.1 and MUSE 8.0.1/CardioSoft V6.51, CardioSoft V6.61, and CardioSoft V6.71 | | |
| RJ45 Wired LAN | ECG Transmission with CSI, DCP Protocol and Shared Directory | | |
| Wireless LAN (WIFI) | ECG Transmission with CSI, DCP Protocol and Shared Directory | | |
| Storage | | | |
| ECG Storage Format | XML format, Hilltop format, PDF storage format | | |
| Storage Capacity | Internal storage of 100 or 200 ECGs | | |
| Accessories | | | |
| ECG Cables/Leadwires | IEC/AHA Value 10LD Patient Cable/Leadwire | | |
| | • 10-lead IEC/AHA Patient trunk cable | | |
| | • 14-lead IEC/AHA Patient trunk cable | | |
| | • IEC/AHA (Nst, Nax) Leadwire set (ECG 10-L w/resist, Banana) | | |
| | IEC/AHA Set of leadwires (4mm connector, 10 leads, defibrillator proof) | | |
| ECG Adapter | IEC/AHA Kit Adapter, 10 set Banana | | |
| | Electrode Prep Pads, CLIP Universal GE 10/Pkg | | |
| Electrodes | ECG Electrode Clamp (Large, 4/set) | | |
| | Baby MAC electrodes | | |
| | Silver Mactrode Plus 1000/CASE | | |
| | Electrode Application System KISS 10-lead without pump | | |
| | Electrode Application System KISS 12-lead without pump | | |
| Other accessories | Electrode Cream 250g bottle, Electrode Spray | | |
| | Country specific power cords | | |
| | • Z-fold Thermal Paper with pre-printed grid and perforation with Queue mark or Queue hole of size 215 mm x 280 mm (Letter)/210 mm x 295mm (A4)/214.2 mm x 279.4 mm (Modified Letter) (150 sheets/pack, 1500 sheets/case) | | |
| | USB data matrix barcode scanner | | |
| | Secure Digital High Capacity Card – 4 GB | | |
| Electrical | | | |
| Power supply | Internal AC/DC or battery operation | | |
| AC/DC operation specifications | Input voltage: 100 to 240 VAC ±10% | | |
| | Input current: Maximum 1.5A in voltage range 115 V to 230 V AC | | |
| | Input frequency: 47 to 63 Hz | | |
| Battery specifications | Battery type: Replaceable and rechargeable, Lithium Ion | | |
| | Battery capacity: 14.4V and 16.8V Max @ 2.2 AH ±10%; 100 single page resting ECG recordings or 3 hours (typical) of continuous monitoring without printing, at a minimum | | |
| | Battery charge time: Approximately 3.5 hours after low battery shut down (with device off) | | |

| Physical specification | |
|-----------------------------|--|
| Height | 200 mm |
| Width | 390 mm |
| Depth | 330 mm |
| Weight | Approx. 5 Kg including battery, without paper |
| Environmental specification | |
| Temperature | Operating: 10°C to 40°C |
| | Transport/storage: -40°C to 70°C |
| Humidity | Operating: 20% to 95% RH non-condensing |
| | Transport/storage: 15% to 95% RH non-condensing |
| Pressure | Operating: 700 to 1060 hPA (Altitude range: 3010.9 to -381.9 meters) |
| | Transport/storage: 500 to 1060 hPA (Altitude range: 5570 to -380 meters) |

Safety and Regulatory

- CE marking for Council Directive 93/42/EEC concerning medical devices
- EN 60601-1 (IEC 60601-1) Medical electrical equipment Part 1: General Requirements for Safety
- IEC 60601-1-1 Medical Electrical Equipment: General Requirements for Safety
- IEC 60601-1-2 General Requirements for Safety Electromagnetic Compatibility
- IEC 60601-1-4 Requirements for Collateral Standard Programmable Electrical Medical Systems
- IEC 60601-2-25 Safety of Electrocardiographs
- IEC 60601-2-51 Safety and performance of ECG recorders
- UL 60601-1: 2006 UL Standard for Safety Medical Electrical Equipment, Part 1: General Requirements for Safety
- CAN/CSA C22.2 No. 601.1 M90
- AAMI EC 11: 1991/(R) 2001/(R) 2007 Diagnostic Electrocardiography Devices
- AAMI EC 13: 2002/(R) 2007 Section 4.2.7 only
- IEC 60601-1-6 General Requirements for Safety Usability

GE Healthcare 8200 West Tower Avenue Milwaukee, WI 53223 U.S.A.

www.gehealthcare.com



 $\hbox{@2013 General Electric Company-All rights reserved}.$

General Electric Company reserves the right to make changes in specifications and features shown herein, or discontinue the product described at any time without notice or obligation.

GE and GE Monogram are trademarks of General Electric Company.

Windows is a registered trademark of Microsoft Corporation.

All other product names and logos are trademarks or registered trademarks of their respective companies.

*Trademark of General Electric Company.

GE Healthcare, a division of General Electric Company.